

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCIA01
Lab Code: ACE **SDG NO.:** Q3215
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114142.D **Time Analyzed:** 15:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.739	4.643	4.843	516.550	500.000	3.3
Aroclor-1016-2	4.757	4.661	4.861	518.190	500.000	3.6
Aroclor-1016-3	4.815	4.718	4.918	504.330	500.000	0.9
Aroclor-1016-4	4.934	4.837	5.037	501.580	500.000	0.3
Aroclor-1016-5	5.190	5.093	5.293	543.000	500.000	8.6
Aroclor-1260-1	6.224	6.129	6.329	555.540	500.000	11.1
Aroclor-1260-2	6.414	6.319	6.519	530.570	500.000	6.1
Aroclor-1260-3	6.779	6.683	6.883	543.010	500.000	8.6
Aroclor-1260-4	7.037	6.941	7.141	612.280	500.000	22.5
Aroclor-1260-5	7.280	7.185	7.385	614.200	500.000	22.8
Decachlorobiphenyl	8.663	8.570	8.770	52.540	50.000	5.1
Tetrachloro-m-xylene	3.658	3.560	3.760	53.330	50.000	6.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCIA01
Lab Code: ACE **SDG NO.:** Q3215
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114142.D **Time Analyzed:** 15:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.719	4.623	4.823	479.090	500.000	-4.2
Aroclor-1016-2	4.737	4.641	4.841	480.850	500.000	-3.8
Aroclor-1016-3	4.911	4.816	5.016	479.280	500.000	-4.1
Aroclor-1016-4	4.954	4.858	5.058	485.070	500.000	-3.0
Aroclor-1016-5	5.165	5.070	5.270	496.440	500.000	-0.7
Aroclor-1260-1	6.191	6.096	6.296	466.710	500.000	-6.7
Aroclor-1260-2	6.379	6.284	6.484	444.900	500.000	-11.0
Aroclor-1260-3	6.530	6.435	6.635	436.540	500.000	-12.7
Aroclor-1260-4	6.998	6.903	7.103	447.410	500.000	-10.5
Aroclor-1260-5	7.240	7.146	7.346	446.550	500.000	-10.7
Decachlorobiphenyl	8.607	8.514	8.714	52.500	50.000	5.0
Tetrachloro-m-xylene	3.649	3.551	3.751	50.570	50.000	1.1