

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCIA01
Lab Code: ACE **SDG NO.:** Q3215
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114229.D **Time Analyzed:** 09:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.741	4.643	4.843	563.070	500.000	12.6
Aroclor-1016-2	4.760	4.661	4.861	560.970	500.000	12.2
Aroclor-1016-3	4.818	4.718	4.918	546.470	500.000	9.3
Aroclor-1016-4	4.936	4.837	5.037	545.680	500.000	9.1
Aroclor-1016-5	5.192	5.093	5.293	581.830	500.000	16.4
Aroclor-1260-1	6.227	6.129	6.329	580.340	500.000	16.1
Aroclor-1260-2	6.417	6.319	6.519	564.850	500.000	13.0
Aroclor-1260-3	6.781	6.683	6.883	577.140	500.000	15.4
Aroclor-1260-4	7.039	6.941	7.141	632.170	500.000	26.4
Aroclor-1260-5	7.282	7.185	7.385	641.060	500.000	28.2
Decachlorobiphenyl	8.666	8.570	8.770	48.790	50.000	-2.4
Tetrachloro-m-xylene	3.660	3.560	3.760	46.630	50.000	-6.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCIA01
Lab Code: ACE **SDG NO.:** Q3215
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114229.D **Time Analyzed:** 09:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.714	4.623	4.823	499.390	500.000	-0.1
Aroclor-1016-2	4.732	4.641	4.841	500.950	500.000	0.2
Aroclor-1016-3	4.907	4.816	5.016	496.350	500.000	-0.7
Aroclor-1016-4	4.950	4.858	5.058	499.730	500.000	-0.1
Aroclor-1016-5	5.161	5.070	5.270	517.420	500.000	3.5
Aroclor-1260-1	6.186	6.096	6.296	496.210	500.000	-0.8
Aroclor-1260-2	6.376	6.284	6.484	475.370	500.000	-4.9
Aroclor-1260-3	6.526	6.435	6.635	466.020	500.000	-6.8
Aroclor-1260-4	6.995	6.903	7.103	488.610	500.000	-2.3
Aroclor-1260-5	7.238	7.146	7.346	483.910	500.000	-3.2
Decachlorobiphenyl	8.605	8.514	8.714	43.980	50.000	-12.0
Tetrachloro-m-xylene	3.644	3.551	3.751	42.260	50.000	-15.5