

**CALIBRATION VERIFICATION SUMMARY**

**Lab Name:** Alliance **Contract:** SCIA01  
**Lab Code:** ACE **SDG NO.:** Q3215  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

**Client Sample No.:** CCAL04 **Date Analyzed:** 10/03/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO114243.D **Time Analyzed:** 15:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.737	4.643	4.843	541.730	500.000	8.3
Aroclor-1016-2	4.755	4.661	4.861	549.220	500.000	9.8
Aroclor-1016-3	4.812	4.718	4.918	525.230	500.000	5.0
Aroclor-1016-4	4.931	4.837	5.037	500.770	500.000	0.2
Aroclor-1016-5	5.187	5.093	5.293	557.030	500.000	11.4
Aroclor-1260-1	6.221	6.129	6.329	541.810	500.000	8.4
Aroclor-1260-2	6.411	6.319	6.519	493.120	500.000	-1.4
Aroclor-1260-3	6.775	6.683	6.883	467.170	500.000	-6.6
Aroclor-1260-4	7.034	6.941	7.141	515.670	500.000	3.1
Aroclor-1260-5	7.277	7.185	7.385	523.950	500.000	4.8
Decachlorobiphenyl	8.659	8.570	8.770	38.850	50.000	-22.3
Tetrachloro-m-xylene	3.656	3.560	3.760	45.710	50.000	-8.6

**CALIBRATION VERIFICATION SUMMARY**

**Lab Name:** Alliance **Contract:** SCIA01  
**Lab Code:** ACE **SDG NO.:** Q3215  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

**Client Sample No.:** CCAL04 **Date Analyzed:** 10/03/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO114243.D **Time Analyzed:** 15:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.716	4.623	4.823	499.220	500.000	-0.2
Aroclor-1016-2	4.734	4.641	4.841	488.590	500.000	-2.3
Aroclor-1016-3	4.909	4.816	5.016	493.130	500.000	-1.4
Aroclor-1016-4	4.951	4.858	5.058	496.700	500.000	-0.7
Aroclor-1016-5	5.163	5.070	5.270	514.750	500.000	3.0
Aroclor-1260-1	6.187	6.096	6.296	454.050	500.000	-9.2
Aroclor-1260-2	6.376	6.284	6.484	432.400	500.000	-13.5
Aroclor-1260-3	6.526	6.435	6.635	415.570	500.000	-16.9
Aroclor-1260-4	6.995	6.903	7.103	418.710	500.000	-16.3
Aroclor-1260-5	7.237	7.146	7.346	411.890	500.000	-17.6
Decachlorobiphenyl	8.603	8.514	8.714	40.300	50.000	-19.4
Tetrachloro-m-xylene	3.647	3.551	3.751	42.450	50.000	-15.1