

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Alliance

Contract: SCIA01

ProjectID: 507 Franklin Ave Nutley

Lab Code: ACE

SDG No.: Q3215

Calibration Sequence : FF090525		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
850	137068275	161257	FF016336.D
340	37986080	111724	FF016337.D
170	21402125	125895	FF016338.D
85	11165009	131353	FF016339.D
1700	176262056	103684	FF016340.D
AVG RF : 126783		% RSD : 17.506	AVG RT : 15.071