

Surrogate Summary

SW-846

SDG No.: Q3215

Client: Sciacca General Contractors, LLC

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169890BL	PB169890BL	Nitrobenzene-d5	100	80.0	80		18	107
		2-Fluorobiphenyl	100	84.3	84		20	109
		Terphenyl-d14	100	106	106	*	10	105
PB169890BS	PB169890BS	Nitrobenzene-d5	100	80.7	81		18	107
		2-Fluorobiphenyl	100	84.0	84		20	109
		Terphenyl-d14	100	99.3	99		10	105
Q3215-01	WASTE	Nitrobenzene-d5	100	43.4	43		18	107
		2-Fluorobiphenyl	100	51.0	51		20	109
		Terphenyl-d14	100	57.7	58		10	105
Q3230-01MS	WALL-PROBE-BRICK-1MS	Nitrobenzene-d5	100	56.6	57		18	107
		2-Fluorobiphenyl	100	50.9	51		20	109
		Terphenyl-d14	100	58.0	58		10	105
Q3230-01MSD	WALL-PROBE-BRICK-1MSD	Nitrobenzene-d5	100	57.8	58		18	107
		2-Fluorobiphenyl	100	51.7	52		20	109
		Terphenyl-d14	100	62.6	63		10	105