

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3215

Analytical Method: SW8270E

Client: Sciaccia General Contractors, LLC

DataFile: BF143825.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q3230-01MSD		Client Sample ID: WALL-PROBE-BRICK-1MSD									
Naphthalene	1200	0	1100	ug/Kg	92	0			51	121	20
Acenaphthylene	1200	0	1300	ug/Kg	108	0			57	125	20
Acenaphthene	1200	0	1200	ug/Kg	100	0			70	121	20
Fluorene	1200	0	1200	ug/Kg	100	0			53	118	20
Phenanthrene	1200	81.0	1200	ug/Kg	93	0			52	128	20
Anthracene	1200	0	1200	ug/Kg	100	8			62	124	20
Fluoranthene	1200	96.8	1200	ug/Kg	92	0			44	125	20
Pyrene	1200	90.5	1400	ug/Kg	109	0			37	122	20
Benzo(a)anthracene	1200	0	1200	ug/Kg	100	0			53	119	20
Chrysene	1200	0	1300	ug/Kg	108	8			57	121	20
Benzo(b)fluoranthene	1200	0	1300	ug/Kg	108	8			52	117	20
Benzo(k)fluoranthene	1200	0	1200	ug/Kg	100	0			57	134	20
Benzo(a)pyrene	1200	0	1300	ug/Kg	108	0			70	142	20
Indeno(1,2,3-cd)pyrene	1200	0	1200	ug/Kg	100	0			40	129	20
Dibenz(a,h)anthracene	1200	0	1200	ug/Kg	100	8			43	123	20
Benzo(g,h,i)perylene	1200	0	1200	ug/Kg	100	0			24	125	20