

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3215

Client ID : SSTDCCC040

Date Analyzed: 09/29/2025

Lab File ID: BF143820.D

Time Analyzed: 11:09

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	142387	6.887	519222	8.17	254612	9.92
UPPER LIMIT	284774	7.387	1038440	8.669	509224	10.422
LOWER LIMIT	71193.5	6.387	259611	7.669	127306	9.422
EPA SAMPLE NO.						
01 PB169890BL	168062	6.89	649975	8.16	352419	9.92
02 PB169890BS	174542	6.89	648417	8.17	329040	9.93
03 WALL-PROBE-BRICK-1MS	137049	6.89	530817	8.17	285401	9.93
04 WALL-PROBE-BRICK-1MSD	130106	6.89	506656	8.17	272537	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.