

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3215

Client ID: SSTDCCC040

Date Analyzed: 09/29/2025

Lab File ID: BG064456.D

Time Analyzed: 12:21

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	131242	17.201	160574	21.431	177303	24.428
UPPER LIMIT	262484	17.701	321148	21.931	354606	24.928
LOWER LIMIT	65621	16.701	80287	20.931	88651.5	23.928
EPA SAMPLE NO.						
01 WASTE	148242	17.20	160128	21.43	172776	24.43

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.