

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG NO.: Q3215
 Lab File ID: VW032267.D Date Analyzed: 09/26/2025
 Instrument ID: MSVOA_W Time Analyzed: 08:43
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	273367	7.97	482464	8.85	450604	11.63
UPPER LIMIT	546734	8.465	964928	9.349	901208	12.129
LOWER LIMIT	136684	7.465	241232	8.349	225302	11.129
EPA SAMPLE NO.						
WASTE	180242	7.96	388056	8.85	386764	11.64
VW0926SBL01	193650	7.96	436371	8.85	462353	11.63
VW0926SBS01	258288	7.96	477257	8.85	426681	11.63
VW0926SBSD01	272106	7.96	497590	8.85	443141	11.63

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.