

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: SCIA01
 Lab Code: ACE SDG NO.: Q3215
 Lab File ID: VW032267.D Date Analyzed: 09/26/2025
 Instrument ID: MSVOA_W Time Analyzed: 08:43
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	228628	13.555				
UPPER LIMIT	457256	14.055				
LOWER LIMIT	114314	13.055				
EPA SAMPLE NO.						
WASTE	183728	13.56				
VW0926SBL01	216659	13.56				
VW0926SBS01	223480	13.56				
VW0926SBSD01	218363	13.55				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.