

Continuing Calibration Report for SequenceID : FC100325AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069919.D

Aliphatic C12-C16	7357331.000	40.000	6.580	9.978	183933.275	156673.343	-17.399
Aliphatic C16-C21	10792891.000	60.000	9.979	13.344	179881.517	179548.834	-0.185
Aliphatic C21-C28	12867433.000	80.000	13.345	17.006	160842.913	180941.740	11.108
Aliphatic C28-C40	15702694.000	120.000	17.007	21.955	130855.783	128677.038	-1.693
Aliphatic EPH	56343510.000	360.000	3.286	21.955	156509.750	152688.452	-2.503
Aliphatic C9-C12	9623161.000	60.000	3.286	6.579	160386.017	133523.254	-20.118