

Continuing Calibration Report for SequenceID : FC100325AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069924.D

Aliphatic C12-C16	7295872.000	40.000	6.580	9.978	182396.800	156673.343	-16.419
Aliphatic C16-C21	10440377.000	60.000	9.979	13.344	174006.283	179548.834	3.087
Aliphatic C21-C28	12242615.000	80.000	13.345	17.006	153032.688	180941.740	15.424
Aliphatic C28-C40	15404400.000	120.000	17.007	21.955	128370.000	128677.038	0.239
Aliphatic EPH	55143867.000	360.000	3.286	21.955	153177.408	152688.452	-0.320
Aliphatic C9-C12	9760603.000	60.000	3.286	6.579	162676.717	133523.254	-21.834