

Surrogate Summary

SDG No.: Q3216

Client: Sciacca General Contractors, LLC

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3216-02	VOC	1,2-Dichloroethane-d4	50	60.2	120		63	155
		Dibromofluoromethane	50	52.0	104		70	134
		Toluene-d8	50	50.9	102		74	123
		4-Bromofluorobenzene	50	46.3	93		17	146
VW0926SBL01	VW0926SBL01	1,2-Dichloroethane-d4	50	63.5	127		63	155
		Dibromofluoromethane	50	53.0	106		70	134
		Toluene-d8	50	53.6	107		74	123
		4-Bromofluorobenzene	50	47.5	95		17	146
VW0926SBS01	VW0926SBS01	1,2-Dichloroethane-d4	50	47.1	94		63	155
		Dibromofluoromethane	50	46.3	93		70	134
		Toluene-d8	50	46.9	94		74	123
		4-Bromofluorobenzene	50	48.4	97		17	146
VW0926SBSD0	VW0926SBSD01	1,2-Dichloroethane-d4	50	44.4	89		63	155
		Dibromofluoromethane	50	44.0	88		70	134
		Toluene-d8	50	46.2	92		74	123
		4-Bromofluorobenzene	50	45.5	91		17	146