

Initial and Continuing Calibration Verification

Client: ATG-GREENVILLE AEC

SDG No.: Q3218

Project: AEC-2025-0013- CSC 7037 White Plains Fisher

RunNo.: LB137422

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.173	50	100	95-105	09/08/2025
Sample ID: CCV1 COD	mg/L	49.161	50	98	95-105	10/03/2025
Sample ID: CCV2 COD	mg/L	48.150	50	96	95-105	10/03/2025
Sample ID: CCV3 COD	mg/L	50.173	50	100	95-105	10/03/2025