

Duplicate Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3219
Project:	AEC-2025-0013- CSC 4027 Worcester	Sample ID:	Q3202-11
Client ID:	SY-3DDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	42.0		42.2		1	0.48		09/30/2025

Duplicate Sample Summary

Client: ATG-GREENVILLE AEC	SDG No.: Q3219
Project: AEC-2025-0013- CSC 4027 Worcester	Sample ID: Q3263-02
Client ID: 266380DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	1.50	U	1.50	U	1	0		10/03/2025

Duplicate Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3219
Project:	AEC-2025-0013- CSC 4027 Worcester	Sample ID:	Q3263-02
Client ID:	266380MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	48.2		47.1		1	2.31		10/03/2025