

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3087

Client ID: SSTD020383

Date Analyzed: 10/03/2025

Lab File ID: BN037998.D

Time Analyzed: 15:27

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1582360	17.21	1299050	21.463	1049600	24.533
UPPER LIMIT	3164720	17.71	2598100	21.963	2099200	25.033
LOWER LIMIT	791180	16.71	649525	20.963	524800	24.033
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	1398120	17.21	1149170	21.46	1137270	24.53

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.