

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: ALLI03
 Lab Code: ACE SDG No.: Q3226
 Instrument ID: MSVOA_X Calibration Date/Time: 10/03/2025 09:35
 Lab File ID: VX047981.D Init. Calib. Date(s): 09/16/2025 09/16/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 09:23 12:01
 GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
2-Chloroethyl Vinyl ether	250.000	291.84	-16.7364	20