

Surrogate Summary

SDG No.: **Q3226**

Client: **Alliance Technical Group, LLC - Newark**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3226-01	STORAGE-BLANK-SOIL-REF-6	1,2-Dichloroethane-d4	50	55.1	110		74	125
		Dibromofluoromethane	50	50.0	100		75	124
		Toluene-d8	50	48.3	97		86	113
		4-Bromofluorobenzene	50	55.7	111		77	121
Q3226-02	STORAGE-BLANK-WATER-REF-5	1,2-Dichloroethane-d4	50	52.3	105		74	125
		Dibromofluoromethane	50	48.8	98		75	124
		Toluene-d8	50	48.4	97		86	113
		4-Bromofluorobenzene	50	55.1	110		77	121
Q3226-03	STORAGE-BLANK-WATER-REF-4	1,2-Dichloroethane-d4	50	52.9	106		74	125
		Dibromofluoromethane	50	49.1	98		75	124
		Toluene-d8	50	47.8	96		86	113
		4-Bromofluorobenzene	50	55.2	110		77	121
Q3226-04	STORAGE-BLANK-SAMPLE-MANAC	1,2-Dichloroethane-d4	50	52.8	106		74	125
		Dibromofluoromethane	50	50.4	101		75	124
		Toluene-d8	50	49.7	99		86	113
		4-Bromofluorobenzene	50	55.4	111		77	121
VX1003WBL01	VX1003WBL01	1,2-Dichloroethane-d4	50	56.6	113		74	125
		Dibromofluoromethane	50	51.0	102		75	124
		Toluene-d8	50	49.4	99		86	113
		4-Bromofluorobenzene	50	57.0	114		77	121
VX1003WBS02	VX1003WBS02	1,2-Dichloroethane-d4	50	53.6	107		74	125
		Dibromofluoromethane	50	55.1	110		75	124
		Toluene-d8	50	53.6	107		86	113
		4-Bromofluorobenzene	50	54.0	108		77	121
VX1003WBSD0	VX1003WBSD02	1,2-Dichloroethane-d4	50	54.0	108		74	125
		Dibromofluoromethane	50	54.8	110		75	124
		Toluene-d8	50	53.1	106		86	113
		4-Bromofluorobenzene	50	56.0	112		77	121