

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ALLI03
 Lab Code: ACE SDG NO.: Q3226
 Lab File ID: VX047981.D Date Analyzed: 10/03/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:35
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	105470	12.00€				
UPPER LIMIT	210940	12.50€				
LOWER LIMIT	52735	11.50€				
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	132559	12.01				
STORAGE-BLANK-WATER-REF-5	139206	12.01				
STORAGE-BLANK-WATER-REF-4	136177	12.01				
STORAGE-BLANK-SAMPLE-MANAGEMENT	136370	12.01				
VX1003WBL01	147497	12.01				
VX1003WBS02	101886	12.01				
VX1003WBSD02	134054	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.