

Initial and Continuing Calibration Blank Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3227

Project: AEC-2025-0013- CSC 4026 Easthampton

RunNo.: LB137422

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB COD	mg/L	< 5.0000	5.0000	U	1.50	10	09/08/2025
Sample ID: CCB1 COD	mg/L	< 5.0000	5.0000	U	1.50	10	10/03/2025
Sample ID: CCB2 COD	mg/L	< 5.0000	5.0000	U	1.50	10	10/03/2025
Sample ID: CCB3 COD	mg/L	< 5.0000	5.0000	U	1.50	10	10/03/2025