

Laboratory Control Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3227
Project:	AEC-2025-0013- CSC 4026 Easthampton	Run No.:	LB137356

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137356BS							
TSS	mg/L	550	532		97	1	90-110	09/30/2025

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3227

Project: AEC-2025-0013- CSC 4026 Easthampton

Run No.: LB137422

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137422BS							
COD	mg/L	50	51.2		102	1	90-110	10/03/2025