

LAB CHRONICLE

OrderID:	Q3229	OrderDate:	9/26/2025 1:42:00 PM
Client:	Europastry	Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ
Contact:	Kevin Carlucci	Location:	D31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3229-01	MH-9-26-25	WATER			09/26/25 12:30			09/26/25
			Oil and Grease	1664A			10/03/25 10:35	
			TPH	1664A			10/03/25 11:30	



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	Europastry	Date Collected:	09/26/25 12:30
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Date Received:	09/26/25
Client Sample ID:	MH-9-26-25	SDG No.:	Q3229
Lab Sample ID:	Q3229-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Oil and Grease	172		1	0.29	5.00	mg/L		10/03/25 10:35	1664A
TPH	45.4		1	0.29	5.00	mg/L		10/03/25 11:30	1664A

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

QC RESULT SUMMARY

Preparation Blank Summary

Client: Europastry

SDG No.: Q3229

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB137413BL							
Oil and Grease	mg/L	< 2.5000	2.5000	U	0.29	5.0	10/03/2025
Sample ID: LB137414BL							
TPH	mg/L	< 2.5000	2.5000	U	0.29	5.0	10/03/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q3229
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3229-01
Client ID:	MH-9-26-25MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	192		172		20.0	1	101		10/03/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q3229
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3229-01
Client ID:	MH-9-26-25MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	192		172		20.0	1	102		10/03/2025

Duplicate Sample Summary

Client:	Europastray	SDG No.:	Q3229
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	LB137414BS
Client ID:	LB137414BSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TPH	mg/L	+/-18	16.7		17.0		1	1.78		10/03/2025

Duplicate Sample Summary

Client:	Europastry	SDG No.:	Q3229
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3229-01
Client ID:	MH-9-26-25MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	192		192		1	0.05		10/03/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3229
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137413

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137413BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	10/03/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3229
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137414

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137414BS							
TPH	mg/L	20.0	16.7		84	1	78-114	10/03/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3229
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137414

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137414BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	10/03/2025