

Initial and Continuing Calibration Verification

Client: ATG-GREENVILLE AEC

SDG No.: Q3234

Project: AEC-2025-0013- CSC 4028 Cumberland

RunNo.: LB137481

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.489	0.50	98	90-110	10/09/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.506	0.50	101	90-110	10/09/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.502	0.50	100	90-110	10/09/2025