

Initial and Continuing Calibration Blank Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3234

Project: AEC-2025-0013- CSC 4028 Cumberland

RunNo.: LB137481

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Phosphorus, Total	mg/L	0.008	0.0250	J	0.0045	0.05	10/09/2025
Sample ID: CCB1 Phosphorus, Total	mg/L	0.007	0.0250	J	0.0045	0.05	10/09/2025
Sample ID: CCB2 Phosphorus, Total	mg/L	0.005	0.0250	J	0.0045	0.05	10/09/2025