

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ATG-GREENVILLE AEC

SDG No.: Q3234

Contract: ATGG01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	1970	2000	98	90 - 110	P	10/01/2025	12:33	LB137389
	Copper	1010	1000	101	90 - 110	P	10/01/2025	12:33	LB137389
	Lead	3860	4000	96	90 - 110	P	10/01/2025	12:33	LB137389

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ATG-GREENVILLE AEC

SDG No.: Q3234

Contract: ATGG01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	5.97	6.0	99	80 - 120	P	10/01/2025	12:41	LB137389
	Copper	21.3	20.0	106	80 - 120	P	10/01/2025	12:41	LB137389
	Lead	11.1	12.0	93	80 - 120	P	10/01/2025	12:41	LB137389

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ATG-GREENVILLE AEC

SDG No.: Q3234

Contract: ATGG01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2490	2500	100	90 - 110	P	10/01/2025	13:10	LB137389
	Copper	1260	1250	100	90 - 110	P	10/01/2025	13:10	LB137389
	Lead	4950	5000	99	90 - 110	P	10/01/2025	13:10	LB137389
CCV02	Cadmium	2500	2500	100	90 - 110	P	10/01/2025	14:02	LB137389
	Copper	1260	1250	100	90 - 110	P	10/01/2025	14:02	LB137389
	Lead	4960	5000	99	90 - 110	P	10/01/2025	14:02	LB137389
CCV03	Cadmium	2540	2500	102	90 - 110	P	10/01/2025	15:00	LB137389
	Copper	1280	1250	102	90 - 110	P	10/01/2025	15:00	LB137389
	Lead	5050	5000	101	90 - 110	P	10/01/2025	15:00	LB137389
CCV04	Cadmium	2520	2500	101	90 - 110	P	10/01/2025	15:51	LB137389
	Copper	1270	1250	101	90 - 110	P	10/01/2025	15:51	LB137389
	Lead	5000	5000	100	90 - 110	P	10/01/2025	15:51	LB137389
CCV05	Cadmium	2500	2500	100	90 - 110	P	10/01/2025	16:42	LB137389
	Copper	1260	1250	100	90 - 110	P	10/01/2025	16:42	LB137389
	Lead	4960	5000	99	90 - 110	P	10/01/2025	16:42	LB137389
CCV06	Cadmium	2530	2500	101	90 - 110	P	10/01/2025	17:32	LB137389
	Copper	1280	1250	102	90 - 110	P	10/01/2025	17:32	LB137389
	Lead	5030	5000	100	90 - 110	P	10/01/2025	17:32	LB137389
CCV07	Cadmium	2500	2500	100	90 - 110	P	10/01/2025	18:28	LB137389
	Copper	1260	1250	101	90 - 110	P	10/01/2025	18:28	LB137389
	Lead	4970	5000	99	90 - 110	P	10/01/2025	18:28	LB137389
CCV08	Cadmium	2450	2500	98	90 - 110	P	10/01/2025	19:19	LB137389
	Copper	1230	1250	99	90 - 110	P	10/01/2025	19:19	LB137389
	Lead	4860	5000	97	90 - 110	P	10/01/2025	19:19	LB137389