

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: ATG-GREENVILLE AEC

SDG No.: Q3234

Contract: ATGG01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	0.61	1.0	61	-5	7	10/01/2025	12:53	LB137389
	Copper	10.9	2.0	543	-18	22	10/01/2025	12:53	LB137389
	Lead	0.44			-12	12	10/01/2025	12:53	LB137389
ICSAB01	Cadmium	1010	972	104	826	1120	10/01/2025	12:57	LB137389
	Copper	503	511	98	434	588	10/01/2025	12:57	LB137389
	Lead	48.5	49.0	99	37	61	10/01/2025	12:57	LB137389