

metals
- 5a -
MATRIX SPIKE SUMMARY

client: ATG-GREENVILLE AEC **level:** low **sdg no.:** Q3234
contract: ATGG01 **lab code:** ACE
matrix: Water **sample id:** Q3237-02 **client id:** RW8-SP303-20250926MS
Percent Solids for Sample: NA **Spiked ID:** Q3237-02MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	88 - 113	95.0		3.00	U	100	95		P
Copper	ug/L	86 - 114	156		4.15	J	150	101		P
Lead	ug/L	86 - 113	470		6.00	U	500	94		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>ATG-GREENVILLE AEC</u>	level: <u>low</u>	sdg no.: <u>Q3234</u>
contract: <u>ATGG01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3237-02</u>	client id: <u>RW8-SP303-20250926MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3237-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	88 - 113	93.4		3.00	U	100	93		P
Copper	ug/L	86 - 114	154		4.15	J	150	100		P
Lead	ug/L	86 - 113	464		6.00	U	500	93		P