

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ATG-GREENVILLE AEC

SDG No.: Q3234

Contract: ATGG01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169912BS							
Cadmium	ug/L	100	93.3		93	88 - 113	P
Copper	ug/L	150	149		99	86 - 114	P
Lead	ug/L	500	464		93	86 - 113	P