

LAB CHRONICLE

OrderID:	Q3235	OrderDate:	9/29/2025 11:26:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	E31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3235-01	RW5-SP100-2025092 6	Water			09/26/25			09/29/25
			SVOC-SIMGroup1	8270-Modified		10/01/25	10/07/25	
Q3235-01DL	RW5-SP100-2025092 6DL	Water			09/26/25			09/29/25
			SVOC-SIMGroup1	8270-Modified		10/01/25	10/07/25	
Q3235-02	RW5-SP201-2025092 6	Water			09/26/25			09/29/25
			SVOC-SIMGroup1	8270-Modified		10/01/25	10/07/25	
Q3235-03	RW5-SP303-2025092 6	Water			09/26/25			09/29/25
			SVOC-SIMGroup1	8270-Modified		10/01/25	10/07/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet SW-846

SDG No.: Q3235

Client: Tetra Tech NUS, Inc.

Sample ID	Client ID		Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID :	RW5-SP100-20250926								
Q3235-01	RW5-SP100-20250926	WATER	1,4-Dioxane	5.600	E	0.07	0.2	0.2	ug/L
			Total Svoc :	5.60					
			Total Concentration:	5.60					
Client ID :	RW5-SP100-20250926DL								
Q3235-01DL	RW5-SP100-20250926DI	WATER	1,4-Dioxane	5.900	D	0.13	0.4	0.4	ug/L
			Total Svoc :	5.90					
			Total Concentration:	5.90					



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/26/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/29/25
Client Sample ID:	RW5-SP100-20250926	SDG No.:	Q3235
Lab Sample ID:	Q3235-01	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN038034.D	1	10/01/25 12:31	10/07/25 06:46	PB169939

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	5.60	E	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.32		30 - 150		80%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.34		30 - 150		86%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.39		55 - 111		97%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.41		53 - 106		103%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.62	*	58 - 132		154%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	9300		7.804			
1146-65-2	Naphthalene-d8	26600		10.605			
15067-26-2	Acenaphthene-d10	14100		14.462			
1517-22-2	Phenanthrene-d10	27800		17.21			
1719-03-5	Chrysene-d12	15900		21.393			
1520-96-3	Perylene-d12	12300		23.712			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/26/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/29/25
Client Sample ID:	RW5-SP100-20250926DL	SDG No.:	Q3235
Lab Sample ID:	Q3235-01DL	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN038057.D	2	10/01/25 12:31	10/07/25 21:46	PB169939

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	5.90	D	0.13	0.40	0.40	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.33		30 - 150		81%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.33		30 - 150		83%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.40		55 - 111		100%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.40		53 - 106		100%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.62	*	58 - 132		155%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	8840		7.804			
1146-65-2	Naphthalene-d8	25100		10.605			
15067-26-2	Acenaphthene-d10	12800		14.452			
1517-22-2	Phenanthrene-d10	23600		17.198			
1719-03-5	Chrysene-d12	12500		21.384			
1520-96-3	Perylene-d12	11000		23.704			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/26/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/29/25
Client Sample ID:	RW5-SP201-20250926	SDG No.:	Q3235
Lab Sample ID:	Q3235-02	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN038035.D	1	10/01/25 12:31	10/07/25 07:22	PB169939

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.33		30 - 150		83%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.34		30 - 150		84%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.42		55 - 111		106%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.43	*	53 - 106		108%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.53		58 - 132		132%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	10100	7.804				
1146-65-2	Naphthalene-d8	27800	10.605				
15067-26-2	Acenaphthene-d10	13100	14.462				
1517-22-2	Phenanthrene-d10	23200	17.211				
1719-03-5	Chrysene-d12	15100	21.393				
1520-96-3	Perylene-d12	14100	23.712				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/26/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/29/25
Client Sample ID:	RW5-SP303-20250926	SDG No.:	Q3235
Lab Sample ID:	Q3235-03	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	980 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN038036.D	1	10/01/25 12:31	10/07/25 07:58	PB169939

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.30		30 - 150		75%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.35		30 - 150		87%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.37		55 - 111		93%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.36		53 - 106		91%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.62	*	58 - 132		154%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	9010	7.804				
1146-65-2	Naphthalene-d8	25100	10.605				
15067-26-2	Acenaphthene-d10	12900	14.462				
1517-22-2	Phenanthrene-d10	25200	17.211				
1719-03-5	Chrysene-d12	14400	21.393				
1520-96-3	Perylene-d12	11200	23.712				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q3235

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169939BL	PB169939BL	2-Methylnaphthalene-d10	0.4	0.32	79		30	150
		Fluoranthene-d10	0.4	0.29	72		30	150
		Nitrobenzene-d5	0.4	0.41	101		55	111
		2-Fluorobiphenyl	0.4	0.40	100		53	106
		Terphenyl-d14	0.4	0.51	127		58	132
PB169939BS	PB169939BS	2-Methylnaphthalene-d10	0.4	0.36	90		30	150
		Fluoranthene-d10	0.4	0.28	70		30	150
		Nitrobenzene-d5	0.4	0.43	107		55	111
		2-Fluorobiphenyl	0.4	0.45	112	*	53	106
		Terphenyl-d14	0.4	0.53	132		58	132
PB169939BSD	PB169939BSD	2-Methylnaphthalene-d10	0.4	0.35	88		30	150
		Fluoranthene-d10	0.4	0.30	75		30	150
		Nitrobenzene-d5	0.4	0.42	105		55	111
		2-Fluorobiphenyl	0.4	0.42	104		53	106
		Terphenyl-d14	0.4	0.53	133	*	58	132
Q3235-01	RW5-SP100-20250926	2-Methylnaphthalene-d10	0.4	0.32	80		30	150
		Fluoranthene-d10	0.4	0.34	86		30	150
		Nitrobenzene-d5	0.4	0.39	97		55	111
		2-Fluorobiphenyl	0.4	0.41	103		53	106
		Terphenyl-d14	0.4	0.62	154	*	58	132
Q3235-01DL	RW5-SP100-20250926DL	2-Methylnaphthalene-d10	0.4	0.33	81		30	150
		Fluoranthene-d10	0.4	0.33	83		30	150
		Nitrobenzene-d5	0.4	0.40	100		55	111
		2-Fluorobiphenyl	0.4	0.40	100		53	106
		Terphenyl-d14	0.4	0.62	155	*	58	132
Q3235-02	RW5-SP201-20250926	2-Methylnaphthalene-d10	0.4	0.33	83		30	150
		Fluoranthene-d10	0.4	0.34	84		30	150
		Nitrobenzene-d5	0.4	0.42	106		55	111
		2-Fluorobiphenyl	0.4	0.43	108	*	53	106
		Terphenyl-d14	0.4	0.53	132		58	132
Q3235-03	RW5-SP303-20250926	2-Methylnaphthalene-d10	0.4	0.30	75		30	150
		Fluoranthene-d10	0.4	0.35	87		30	150
		Nitrobenzene-d5	0.4	0.37	93		55	111
		2-Fluorobiphenyl	0.4	0.36	91		53	106
		Terphenyl-d14	0.4	0.62	154	*	58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3235

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN038043.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169939BS	1,4-Dioxane	0.4	0.35	ug/L	88				70	130	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3235

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN038044.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169939BSD	1,4-Dioxane	0.4	0.35	ug/L	88	0			70	130	20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

Client ID

PB169939BL

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q3235

Lab File ID: BN038042.D

Lab Sample ID: PB169939BL

Instrument ID: BNA_N

Date Extracted: 10/01/2025

Matrix: (soil/water) Water

Date Analyzed: 10/07/2025

Level: (low/med) LOW

Time Analyzed: 12:39

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB169939BS	PB169939BS	BN038043.D	10/07/2025
PB169939BSD	PB169939BSD	BN038044.D	10/07/2025
RW5-SP100-20250926	Q3235-01	BN038034.D	10/07/2025
RW5-SP201-20250926	Q3235-02	BN038035.D	10/07/2025
RW5-SP303-20250926	Q3235-03	BN038036.D	10/07/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037860.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3235
DFTPP Injection Date: 09/23/2025
DFTPP Injection Time: 11:33

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	4.2
441	Present, but less than mass 443	72.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17.7 (20.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN037861.D	09/23/2025	12:13
SSTDICC0.2	SSTDICC0.2	BN037862.D	09/23/2025	12:49
SSTDICCC0.4	SSTDICCC0.4	BN037863.D	09/23/2025	13:25
SSTDICC0.8	SSTDICC0.8	BN037864.D	09/23/2025	14:02
SSTDICC1.6	SSTDICC1.6	BN037865.D	09/23/2025	14:38
SSTDICC3.2	SSTDICC3.2	BN037866.D	09/23/2025	15:15
SSTDICC5.0	SSTDICC5.0	BN037867.D	09/23/2025	15:51

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN038021.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3235
DFTPP Injection Date: 10/06/2025
DFTPP Injection Time: 22:13

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	81.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.7 (21.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN038022.D	10/06/2025	23:32
RW5-SP100-20250926	Q3235-01	BN038034.D	10/07/2025	06:46
RW5-SP201-20250926	Q3235-02	BN038035.D	10/07/2025	07:22
RW5-SP303-20250926	Q3235-03	BN038036.D	10/07/2025	07:58
SSTDCCC0.4EC	SSTDCCC0.4	BN038039.D	10/07/2025	09:47

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN038040.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3235
DFTPP Injection Date: 10/07/2025
DFTPP Injection Time: 10:44

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.7) 1
69	Mass 69 relative abundance	39.2
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	76.8
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.3 (18.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN038041.D	10/07/2025	12:03
PB169939BL	PB169939BL	BN038042.D	10/07/2025	12:39
PB169939BS	PB169939BS	BN038043.D	10/07/2025	13:16
PB169939BSD	PB169939BSD	BN038044.D	10/07/2025	13:52
RW5-SP100-20250926DL	Q3235-01DL	BN038057.D	10/07/2025	21:46
SSTDCCC0.4EC	SSTDCCC0.4	BN038058.D	10/07/2025	22:22

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3235

Client ID : SSTDCCC0.4

Date Analyzed: 10/06/2025

Lab File ID: BN038022.D

Time Analyzed: 23:32

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	8262	7.804	22968	10.61	12134	14.46
UPPER LIMIT	16524	8.304	45936	11.105	24268	14.962
LOWER LIMIT	4131	7.304	11484	10.105	6067	13.962
EPA SAMPLE NO.						
01 RW5-SP100-20250926	9297	7.80	26599	10.61	14078	14.46
02 RW5-SP201-20250926	10097	7.80	27762	10.61	13082	14.46
03 RW5-SP303-20250926	9005	7.80	25127	10.61	12851	14.46

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3235

Client ID: SSTDCCC0.4

Date Analyzed: 10/06/2025

Lab File ID: BN038022.D

Time Analyzed: 23:32

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	22733	17.21	13562	21.384	10530	23.703
UPPER LIMIT	45466	17.71	27124	21.884	21060	24.203
LOWER LIMIT	11366.5	16.71	6781	20.884	5265	23.203
EPA SAMPLE NO.						
01 RW5-SP100-20250926	27781	17.21	15883	21.39	12313	23.71
02 RW5-SP201-20250926	23156	17.21	15070	21.39	14086	23.71
03 RW5-SP303-20250926	25186	17.21	14411	21.39	11244	23.71

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3235

Client ID : SSTDCCC0.4

Date Analyzed: 10/07/2025

Lab File ID: BN038041.D

Time Analyzed: 12:03

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	8417	7.804	24486	10.61	13140	14.46
UPPER LIMIT	16834	8.304	48972	11.105	26280	14.962
LOWER LIMIT	4208.5	7.304	12243	10.105	6570	13.962
EPA SAMPLE NO.						
01 PB169939BL	8228	7.80	21819	10.61	10027	14.46
02 PB169939BS	8722	7.80	23265	10.61	10633	14.45
03 PB169939BSD	8275	7.80	22197	10.61	10577	14.45
04 RW5-SP100-20250926DL	8836	7.80	25148	10.61	12831	14.45

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3235

Client ID: SSTDCCC0.4

Date Analyzed: 10/07/2025

Lab File ID: BN038041.D

Time Analyzed: 12:03

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	24897	17.21	15245	21.393	11156	23.709
UPPER LIMIT	49794	17.71	30490	21.893	22312	24.209
LOWER LIMIT	12448.5	16.71	7622.5	20.893	5578	23.209
EPA SAMPLE NO.						
01 PB169939BL	17932	17.21	9516	21.39	9042	23.71
02 PB169939BS	17722	17.21	8733	21.39	8119	23.71
03 PB169939BSD	20035	17.20	10845	21.39	8550	23.71
04 RW5-SP100-20250926DL	23567	17.20	12529	21.38	11024	23.70

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB169939BL		SDG No.:	Q3235
Lab Sample ID:	PB169939BL		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN038042.D	1	10/01/25 12:31	10/07/25 12:39	PB169939

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.32		30 - 150		79%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.29		30 - 150		72%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.41		55 - 111		101%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.40		53 - 106		100%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.51		58 - 132		127%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	8230	7.804				
1146-65-2	Naphthalene-d8	21800	10.605				
15067-26-2	Acenaphthene-d10	10000	14.463				
1517-22-2	Phenanthrene-d10	17900	17.211				
1719-03-5	Chrysene-d12	9520	21.393				
1520-96-3	Perylene-d12	9040	23.709				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:		
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:		
Client Sample ID:	PB169939BS		SDG No.:	Q3235	
Lab Sample ID:	PB169939BS		Matrix:	Water	
Analytical Method:	SW8270ESIM		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN038043.D	1	10/01/25 12:31	10/07/25 13:16	PB169939

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.35		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.36		30 - 150		90%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.28		30 - 150		70%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.43		55 - 111		107%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.45	*	53 - 106		112%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.53		58 - 132		132%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	8720	7.804				
1146-65-2	Naphthalene-d8	23300	10.605				
15067-26-2	Acenaphthene-d10	10600	14.452				
1517-22-2	Phenanthrene-d10	17700	17.21				
1719-03-5	Chrysene-d12	8730	21.393				
1520-96-3	Perylene-d12	8120	23.709				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB169939BSD		SDG No.:	Q3235
Lab Sample ID:	PB169939BSD		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN038044.D	1	10/01/25 12:31	10/07/25 13:52	PB169939

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.35		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.35		30 - 150		88%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.30		30 - 150		75%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.42		55 - 111		105%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.42		53 - 106		104%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.53	*	58 - 132		133%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	8280	7.804				
1146-65-2	Naphthalene-d8	22200	10.605				
15067-26-2	Acenaphthene-d10	10600	14.452				
1517-22-2	Phenanthrene-d10	20000	17.198				
1719-03-5	Chrysene-d12	10800	21.393				
1520-96-3	Perylene-d12	8550	23.709				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN092325.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Sep 24 11:00:01 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN037861.D 0.2 =BN037862.D 0.4 =BN037863.D 0.8 =BN037864.D 1.6 =BN037865.D 3.2 =BN037866.D 5 =BN037867.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.413	0.410	0.403	0.425	0.410	0.376	0.406	4.04	
3)	n-Nitrosodimet...	0.541	0.517	0.526	0.561	0.549	0.547	0.540	3.03	
4) S	2-Fluorophenol	0.944	0.921	0.856	0.839	0.908	0.923	0.998	0.913	5.84
5) S	Phenol-d6	1.190	1.200	1.118	1.088	1.211	1.236	1.353	1.199	7.15
6)	bis(2-Chloroet...	1.133	1.096	1.082	1.066	1.150	1.124	1.143	1.113	2.89
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.299	0.289	0.289	0.290	0.310	0.320	0.338	0.305	6.14
9)	Naphthalene	1.105	1.089	1.048	1.051	1.123	1.137	1.160	1.102	3.84
10)	Hexachlorobuta...	0.198	0.184	0.182	0.183	0.192	0.188	0.188	0.188	2.94
11) SURR2	Methylnaphth...	0.515	0.518	0.499	0.504	0.553	0.587	0.714	0.556	13.74
12)	2-Methylnaphth...	0.678	0.681	0.668	0.682	0.754	0.775	0.802	0.720	7.68
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.148	0.142	0.135	0.139	0.163	0.184	0.152	12.15	
15) S	2-Fluorobiphenyl	1.660	1.631	1.559	1.547	1.654	1.650	1.762	1.638	4.36
16)	Acenaphthylene	1.698	1.721	1.663	1.680	1.907	1.951	2.094	1.816	9.23
17)	Acenaphthene	1.262	1.230	1.198	1.208	1.351	1.371	1.437	1.294	7.16
18)	Fluorene	1.455	1.455	1.452	1.466	1.664	1.717	1.748	1.565	8.77
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.041	0.044	0.046	0.058	0.071	0.084	0.057	29.75	
21)	4-Bromophenyl-...	0.252	0.247	0.244	0.244	0.272	0.275	0.291	0.261	7.07
22)	Hexachlorobenzene	0.317	0.312	0.305	0.300	0.324	0.323	0.332	0.316	3.53
23)	Atrazine	0.180	0.181	0.172	0.178	0.205	0.224	0.249	0.199	14.70
24)	Pentachlorophenol	0.101	0.092	0.097	0.116	0.136	0.108	16.50		
25)	Phenanthrene	1.301	1.271	1.230	1.224	1.358	1.379	1.452	1.316	6.37
26)	Anthracene	1.034	1.035	1.023	1.057	1.209	1.275	1.361	1.142	12.10
27) SURR	Fluoranthene-d10	0.880	0.918	0.887	0.900	1.004	1.085	1.370	1.006	17.59
28)	Fluoranthene	1.278	1.315	1.311	1.358	1.555	1.617	1.713	1.450	12.07
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.576	1.600	1.521	1.476	1.607	1.595	1.679	1.579	4.12
31) S	Terphenyl-d14	0.799	0.791	0.753	0.735	0.798	0.809	0.893	0.797	6.34
32)	Benzo(a)anthra...	1.376	1.359	1.294	1.285	1.423	1.483	1.567	1.398	7.27
33)	Chrysene	1.508	1.526	1.462	1.446	1.559	1.512	1.569	1.512	3.03
34)	Bis(2-ethylhex...	0.592	0.487	0.476	0.518	0.569	0.676	0.553	13.57	
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN092325.M

36)	Indeno(1,2,3-c...	1.715	1.672	1.658	1.707	1.887	1.996	2.159	1.828	10.52
37)	Benzo(b)fluora...	1.599	1.606	1.641	1.590	1.810	1.848	1.981	1.725	8.95
38)	Benzo(k)fluora...	1.620	1.586	1.597	1.628	1.875	1.917	1.926	1.736	9.26
39) C	Benzo(a)pyrene	1.254	1.296	1.231	1.254	1.413	1.492	1.610	1.364	10.61
40)	Dibenzo(a,h)an...	1.285	1.283	1.303	1.333	1.495	1.587	1.705	1.427	11.87
41)	Benzo(g,h,i)pe...	1.407	1.381	1.404	1.397	1.521	1.613	1.773	1.499	9.85

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG No.: Q3235
Instrument ID: BNA_N Calibration Date/Time: 10/06/2025 23:32
Lab File ID: BN038022.D Init. Calib. Date(s): 09/23/2025 09/23/2025
EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 12:13 15:51
GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.556	0.504		-9.4	20.0
Fluoranthene-d10	1.006	0.813		-19.2	20.0
2-Fluorophenol	0.913	0.764		-16.3	20.0
Phenol-d6	1.199	1.067		-11.0	20.0
Nitrobenzene-d5	0.305	0.290		-4.9	20.0
2-Fluorobiphenyl	1.638	1.579		-3.6	20.0
2,4,6-Tribromophenol	0.152	0.125		-17.8	20.0
Terphenyl-d14	0.797	0.905		13.6	20.0
1,4-Dioxane	0.406	0.372		-8.4	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3235</u>
Instrument ID:	<u>BNA_N</u>	Calibration Date/Time:	<u>10/07/2025 09:47</u>
Lab File ID:	<u>BN038039.D</u>	Init. Calib. Date(s):	<u>09/23/2025 09/23/2025</u>
EPA Sample No.:	<u>SSTDCCC0.4EC</u>	Init. Calib. Time(s):	<u>12:13 15:51</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.556	0.515		-7.4	50.0
Fluoranthene-d10	1.006	0.820		-18.5	50.0
2-Fluorophenol	0.913	0.850		-6.9	50.0
Phenol-d6	1.199	1.108		-7.6	50.0
Nitrobenzene-d5	0.305	0.291		-4.6	50.0
2-Fluorobiphenyl	1.638	1.521		-7.1	50.0
2,4,6-Tribromophenol	0.152	0.117		-23.0	50.0
Terphenyl-d14	0.797	0.944		18.4	50.0
1,4-Dioxane	0.406	0.426		4.9	50.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3235</u>
Instrument ID:	<u>BNA_N</u>	Calibration Date/Time:	<u>10/07/2025 12:03</u>
Lab File ID:	<u>BN038041.D</u>	Init. Calib. Date(s):	<u>09/23/2025 09/23/2025</u>
EPA Sample No.:	<u>SSTDCCC0.4</u>	Init. Calib. Time(s):	<u>12:13 15:51</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.556	0.513		-7.7	20.0
Fluoranthene-d10	1.006	0.830		-17.5	20.0
2-Fluorophenol	0.913	0.853		-6.6	20.0
Phenol-d6	1.199	1.108		-7.6	20.0
Nitrobenzene-d5	0.305	0.286		-6.2	20.0
2-Fluorobiphenyl	1.638	1.517		-7.4	20.0
2,4,6-Tribromophenol	0.152	0.120		-21.1	20.0
Terphenyl-d14	0.797	0.916		14.9	20.0
1,4-Dioxane	0.406	0.418		3.0	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3235</u>
Instrument ID:	<u>BNA_N</u>	Calibration Date/Time:	<u>10/07/2025 22:22</u>
Lab File ID:	<u>BN038058.D</u>	Init. Calib. Date(s):	<u>09/23/2025 09/23/2025</u>
EPA Sample No.:	<u>SSTDCCC0.4EC</u>	Init. Calib. Time(s):	<u>12:13 15:51</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.556	0.505		-9.2	50.0
Fluoranthene-d10	1.006	0.755		-25.0	50.0
2-Fluorophenol	0.913	0.878		-3.8	50.0
Phenol-d6	1.199	1.126		-6.1	50.0
Nitrobenzene-d5	0.305	0.295		-3.3	50.0
2-Fluorobiphenyl	1.638	1.615		-1.4	50.0
2,4,6-Tribromophenol	0.152	0.112		-26.3	50.0
Terphenyl-d14	0.797	0.862		8.2	50.0
1,4-Dioxane	0.406	0.424		4.4	50.0

All other compounds must meet a minimum RRF of 0.010.