

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN092325

Review By	Rahul	Review On	9/24/2025 8:51:58 AM		
Supervise By	Jagrut	Supervise On	9/24/2025 9:44:29 AM		
SubDirectory	BN092325	HP Acquire Method	BNA_N	HP Processing Method	bn092325

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6842, SP6843, SP6844, SP6845, SP6846, SP6847, SP6848
CCC	SP6846
Internal Standard/PEM	SP6830, 1ul/100ul sample
ICV/I.BLK	SP6854
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037860.D	23 Sep 2025 11:33		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN037861.D	23 Sep 2025 12:13		RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN037862.D	23 Sep 2025 12:49		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN037863.D	23 Sep 2025 13:25	Compound #20 kept on QR	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN037864.D	23 Sep 2025 14:02		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN037865.D	23 Sep 2025 14:38		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN037866.D	23 Sep 2025 15:15		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN037867.D	23 Sep 2025 15:51	Compound#14,24 removed from 5.0 ppm	RC/JU	Ok
9	SSTDICV0.4	ICVBN092325	BN037868.D	23 Sep 2025 16:28		RC/JU	Ok
10	PB169743BL	PB169743BL	BN037869.D	23 Sep 2025 17:42	END CCC missing, Analyzed for Contamination check.	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN100625

Review By	rahul	Review On	10/7/2025 2:42:38 PM
Supervise By	Jagrut	Supervise On	10/7/2025 5:49:36 PM
SubDirectory	BN100625	HP Acquire Method	BNA_N
		HP Processing Method	bn092325

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848
CCC	SP6846
Internal Standard/PEM	SP6830,1ul/100ul sample
ICV/I.BLK	SP6854
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038003.D	06 Oct 2025 10:27		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN038004.D	06 Oct 2025 11:06		RC/JU	Ok
3	PB169903BL	PB169903BL	BN038005.D	06 Oct 2025 11:55		RC/JU	Ok
4	Q3194-02	RW11-MW01S-202509	BN038006.D	06 Oct 2025 12:31		RC/JU	Ok
5	Q3194-03MS	RW11-MW01S-202509	BN038007.D	06 Oct 2025 13:44		RC/JU	Ok
6	Q3194-04MSD	RW11-MW01S-202509	BN038008.D	06 Oct 2025 14:20		RC/JU	Ok,M
7	Q3194-01	TT189D1-20250922	BN038009.D	06 Oct 2025 14:56		RC/JU	Ok
8	Q3194-05	TT150S1-20250922	BN038010.D	06 Oct 2025 15:33		RC/JU	Ok
9	Q3194-06	RW8-MW01D2-202509	BN038011.D	06 Oct 2025 16:09		RC/JU	Ok
10	Q3194-07	RW8-MW01D3-202509	BN038012.D	06 Oct 2025 16:46		RC/JU	Ok
11	Q3194-08	TT188S1-20250922	BN038013.D	06 Oct 2025 17:22		RC/JU	Ok
12	Q3194-09	TT192D2-20250922	BN038014.D	06 Oct 2025 17:59		RC/JU	Ok
13	Q3194-10	TT192D1-20250922	BN038015.D	06 Oct 2025 18:35		RC/JU	Ok
14	Q3194-11	DUP10-20250922	BN038016.D	06 Oct 2025 19:12		RC/JU	Ok
15	Q3194-12	RW11-MW01I-2025092	BN038017.D	06 Oct 2025 19:48		RC/JU	Ok
16	Q3194-13	TT163S1-20250924	BN038018.D	06 Oct 2025 20:24		RC/JU	Ok
17	PB169903BS	PB169903BS	BN038019.D	06 Oct 2025 21:01		RC/JU	Ok
18	SSTDCCC0.4	SSTDCCC0.4EC	BN038020.D	06 Oct 2025 21:37		RC/JU	Ok,M

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN100625

Review By	rahul	Review On	10/7/2025 2:42:38 PM		
Supervise By	Jagrut	Supervise On	10/7/2025 5:49:36 PM		
SubDirectory	BN100625	HP Acquire Method	BNA_N	HP Processing Method	bn092325
STD. NAME		STD REF.#			
Tune/Reschk		SP6875			
Initial Calibration Stds		SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848			
CCC		SP6846			
Internal Standard/PEM		SP6830,1ul/100ul sample			
ICV/I.BLK		SP6854			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	DFTPP	DFTPP	BN038021.D	06 Oct 2025 22:13		RC/JU	Ok
20	SSTDCCC0.4	SSTDCCC0.4	BN038022.D	06 Oct 2025 23:32		RC/JU	Ok,M
21	PB169904BL	PB169904BL	BN038023.D	07 Oct 2025 00:08		RC/JU	Ok
22	Q3194-14	RE139D1-20250924	BN038024.D	07 Oct 2025 00:44		RC/JU	Ok
23	Q3194-15	EB05-20250924	BN038025.D	07 Oct 2025 01:21		RC/JU	Ok
24	Q3194-16	EB06-20250924	BN038026.D	07 Oct 2025 01:57		RC/JU	Ok
25	Q3194-17	FB01-20250924	BN038027.D	07 Oct 2025 02:33		RC/JU	Ok
26	Q3213-02	VPB184-HYD-20250924	BN038028.D	07 Oct 2025 03:09	Surrogate Fail	RC/JU	Ok
27	Q3213-03	BP-VPB-184-GW-340-3	BN038029.D	07 Oct 2025 03:46		RC/JU	Ok
28	Q3213-05	BP-VPB-184-GW-380-3	BN038030.D	07 Oct 2025 04:22		RC/JU	Ok
29	Q3213-06	BP-VPB-184-DUP-20250924	BN038031.D	07 Oct 2025 04:58		RC/JU	Ok
30	Q3213-08	BP-VPB-184-GW-420-4	BN038032.D	07 Oct 2025 05:34		RC/JU	Ok
31	Q3213-10	BP-VPB-184-EB-20250924	BN038033.D	07 Oct 2025 06:10		RC/JU	Ok
32	Q3235-01	RW5-SP100-20250926	BN038034.D	07 Oct 2025 06:46	Need 2X Dilution	RC/JU	Dilution
33	Q3235-02	RW5-SP201-20250926	BN038035.D	07 Oct 2025 07:22		RC/JU	Ok
34	Q3235-03	RW5-SP303-20250926	BN038036.D	07 Oct 2025 07:58		RC/JU	Ok
35	PB169904BS	PB169904BS	BN038037.D	07 Oct 2025 08:35		RC/JU	Ok
36	PB169904BSD	PB169904BSD	BN038038.D	07 Oct 2025 09:11		RC/JU	Ok
37	SSTDCCC0.4	SSTDCCC0.4EC	BN038039.D	07 Oct 2025 09:47		RC/JU	Ok

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN100625

Review By	rahul	Review On	10/7/2025 2:42:38 PM		
Supervise By	Jagrut	Supervise On	10/7/2025 5:49:36 PM		
SubDirectory	BN100625	HP Acquire Method	BNA_N	HP Processing Method	bn092325
STD. NAME	STD REF.#				
Tune/Reschk	SP6875				
Initial Calibration Stds	SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848				
CCC	SP6846				
Internal Standard/PEM	SP6830,1ul/100ul sample				
ICV/I.BLK	SP6854				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN100725

Review By	Rahul	Review On	10/8/2025 3:02:32 PM
Supervise By	Jagrut	Supervise On	10/8/2025 3:03:47 PM
SubDirectory	BN100725	HP Acquire Method	BNA_N, 8270_HP Processing Method bn092325

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6842, SP6843, SP6844, SP6845, SP6846, SP6847, SP6848
CCC	SP6846
Internal Standard/PEM	SP6830, 1ul/100ul sample
ICV/I.BLK	SP6854
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038040.D	07 Oct 2025 10:44		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN038041.D	07 Oct 2025 12:03		RC/JU	Ok
3	PB169939BL	PB169939BL	BN038042.D	07 Oct 2025 12:39		RC/JU	Ok
4	PB169939BS	PB169939BS	BN038043.D	07 Oct 2025 13:16		RC/JU	Ok
5	PB169939BSD	PB169939BSD	BN038044.D	07 Oct 2025 13:52		RC/JU	Ok
6	Q3248-04	BP-VPB-184-GW-600-6	BN038045.D	07 Oct 2025 14:28		RC/JU	Ok
7	Q3248-05	BP-VPB-184-GW-620-6	BN038046.D	07 Oct 2025 15:05		RC/JU	Ok
8	Q3248-07	BP-VPB-184-GW-660-6	BN038047.D	07 Oct 2025 15:41		RC/JU	Ok
9	Q3236-01	RW7-SP100-20250926	BN038048.D	07 Oct 2025 16:18		RC/JU	Ok
10	Q3236-02	RW7-SP201-20250926	BN038049.D	07 Oct 2025 16:54		RC/JU	Ok
11	Q3236-03	RW7-SP302-20250926	BN038050.D	07 Oct 2025 17:31		RC/JU	Ok
12	Q3236-04	RW7-SP303-20250926	BN038051.D	07 Oct 2025 18:07		RC/JU	Ok
13	Q3237-01	RW8-SP100-20250926	BN038052.D	07 Oct 2025 18:44		RC/JU	Ok
14	Q3237-02	RW8-SP303-20250926	BN038053.D	07 Oct 2025 19:20		RC/JU	Ok
15	Q3246-02	BP-VPB-184-GW-460-4	BN038054.D	07 Oct 2025 19:57		RC/JU	Ok
16	Q3246-04	BP-VPB-184-GW-500-5	BN038055.D	07 Oct 2025 20:33		RC/JU	Ok
17	Q3246-07	BP-VPB-184-GW-540-5	BN038056.D	07 Oct 2025 21:10		RC/JU	Ok
18	Q3235-01DL	RW5-SP100-20250926	BN038057.D	07 Oct 2025 21:46		RC/JU	Ok

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN100725

Review By	Rahul	Review On	10/8/2025 3:02:32 PM
Supervise By	Jagrut	Supervise On	10/8/2025 3:03:47 PM
SubDirectory	BN100725	HP Acquire Method	BNA_N, 8270_HP Processing Method bn092325
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848		
CCC	SP6846		
Internal Standard/PEM	SP6830,1ul/100ul sample		
ICV/I.BLK	SP6854		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	SSTDCCC0.4	SSTDCCC0.4EC	BN038058.D	07 Oct 2025 22:22		RC/JU	Ok,M
20	DFTPP	DFTPP	BN038059.D	07 Oct 2025 22:58		RC/JU	Ok
21	SSTDCCC0.4	SSTDCCC0.4	BN038060.D	08 Oct 2025 00:17	CCAL failed for Fluoranthene-d10	RC/JU	Ok,M
22	PB169992BL	PB169992BL	BN038061.D	08 Oct 2025 00:54	CCAL failed	RC/JU	Not Ok
23	PB169992BS	PB169992BS	BN038062.D	08 Oct 2025 01:30	CCAL failed	RC/JU	Not Ok
24	PB169992BSD	PB169992BSD	BN038063.D	08 Oct 2025 02:06	CCAL failed	RC/JU	Not Ok
25	SSTDCCC0.4	SSTDCCC0.4EC	BN038064.D	08 Oct 2025 02:43		RC/JU	Ok

M : Manual Integration