

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3235

Client ID : SSTDCCC0.4

Date Analyzed: 10/07/2025

Lab File ID: BN038041.D

Time Analyzed: 12:03

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	8417	7.804	24486	10.61	13140	14.46
UPPER LIMIT	16834	8.304	48972	11.105	26280	14.962
LOWER LIMIT	4208.5	7.304	12243	10.105	6570	13.962
EPA SAMPLE NO.						
01 PB169939BL	8228	7.80	21819	10.61	10027	14.46
02 PB169939BS	8722	7.80	23265	10.61	10633	14.45
03 PB169939BSD	8275	7.80	22197	10.61	10577	14.45
04 RW5-SP100-20250926DL	8836	7.80	25148	10.61	12831	14.45

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.