

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3237

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	3860	4000	96	90 - 110	P	10/01/2025	12:33	LB137389



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3237

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	107	100	107	80 - 120	P	10/01/2025	12:41	LB137389

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q3237

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	4920	5000	98	90 - 110	P	10/01/2025	13:10	LB137389
CCV02	Iron	4910	5000	98	90 - 110	P	10/01/2025	14:02	LB137389
CCV03	Iron	4980	5000	100	90 - 110	P	10/01/2025	15:00	LB137389
CCV04	Iron	4960	5000	99	90 - 110	P	10/01/2025	15:51	LB137389
CCV05	Iron	4990	5000	100	90 - 110	P	10/01/2025	16:42	LB137389
CCV06	Iron	4940	5000	99	90 - 110	P	10/01/2025	17:32	LB137389
CCV07	Iron	5110	5000	102	90 - 110	P	10/01/2025	18:28	LB137389
CCV08	Iron	4870	5000	97	90 - 110	P	10/01/2025	19:19	LB137389