

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN092325

Review By	Rahul	Review On	9/24/2025 8:51:58 AM		
Supervise By	Jagrut	Supervise On	9/24/2025 9:44:29 AM		
SubDirectory	BN092325	HP Acquire Method	BNA_N	HP Processing Method	bn092325

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848
CCC	SP6846
Internal Standard/PEM	SP6830,1ul/100ul sample
ICV/I.BLK	SP6854
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037860.D	23 Sep 2025 11:33		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN037861.D	23 Sep 2025 12:13		RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN037862.D	23 Sep 2025 12:49		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN037863.D	23 Sep 2025 13:25	Compound #20 kept on QR	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN037864.D	23 Sep 2025 14:02		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN037865.D	23 Sep 2025 14:38		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN037866.D	23 Sep 2025 15:15		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN037867.D	23 Sep 2025 15:51	Compound#14,24 removed from 5.0 ppm	RC/JU	Ok
9	SSTDICV0.4	ICVBN092325	BN037868.D	23 Sep 2025 16:28		RC/JU	Ok
10	PB169743BL	PB169743BL	BN037869.D	23 Sep 2025 17:42	END CCC missing, Analyzed for Contamination check.	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN100725

Review By	Rahul	Review On	10/8/2025 3:02:32 PM
Supervise By	Jagrut	Supervise On	10/8/2025 3:03:47 PM
SubDirectory	BN100725	HP Acquire Method	BNA_N, 8270_HP Processing Method bn092325

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6875 SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6846 SP6830,1ul/100ul sample SP6854

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038040.D	07 Oct 2025 10:44		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN038041.D	07 Oct 2025 12:03		RC/JU	Ok
3	PB169939BL	PB169939BL	BN038042.D	07 Oct 2025 12:39		RC/JU	Ok
4	PB169939BS	PB169939BS	BN038043.D	07 Oct 2025 13:16		RC/JU	Ok
5	PB169939BSD	PB169939BSD	BN038044.D	07 Oct 2025 13:52		RC/JU	Ok
6	Q3248-04	BP-VPB-184-GW-600-6	BN038045.D	07 Oct 2025 14:28		RC/JU	Ok
7	Q3248-05	BP-VPB-184-GW-620-6	BN038046.D	07 Oct 2025 15:05		RC/JU	Ok
8	Q3248-07	BP-VPB-184-GW-660-6	BN038047.D	07 Oct 2025 15:41		RC/JU	Ok
9	Q3236-01	RW7-SP100-20250926	BN038048.D	07 Oct 2025 16:18		RC/JU	Ok
10	Q3236-02	RW7-SP201-20250926	BN038049.D	07 Oct 2025 16:54		RC/JU	Ok
11	Q3236-03	RW7-SP302-20250926	BN038050.D	07 Oct 2025 17:31		RC/JU	Ok
12	Q3236-04	RW7-SP303-20250926	BN038051.D	07 Oct 2025 18:07		RC/JU	Ok
13	Q3237-01	RW8-SP100-20250926	BN038052.D	07 Oct 2025 18:44		RC/JU	Ok
14	Q3237-02	RW8-SP303-20250926	BN038053.D	07 Oct 2025 19:20		RC/JU	Ok
15	Q3246-02	BP-VPB-184-GW-460-4	BN038054.D	07 Oct 2025 19:57		RC/JU	Ok
16	Q3246-04	BP-VPB-184-GW-500-5	BN038055.D	07 Oct 2025 20:33		RC/JU	Ok
17	Q3246-07	BP-VPB-184-GW-540-5	BN038056.D	07 Oct 2025 21:10		RC/JU	Ok
18	Q3235-01DL	RW5-SP100-20250926	BN038057.D	07 Oct 2025 21:46		RC/JU	Ok

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN100725

Review By	Rahul	Review On	10/8/2025 3:02:32 PM
Supervise By	Jagrut	Supervise On	10/8/2025 3:03:47 PM
SubDirectory	BN100725	HP Acquire Method	BNA_N, 8270_HP Processing Method bn092325
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848		
CCC	SP6846		
Internal Standard/PEM	SP6830,1ul/100ul sample		
ICV/I.BLK	SP6854		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	SSTDCCC0.4	SSTDCCC0.4EC	BN038058.D	07 Oct 2025 22:22		RC/JU	Ok,M
20	DFTPP	DFTPP	BN038059.D	07 Oct 2025 22:58		RC/JU	Ok
21	SSTDCCC0.4	SSTDCCC0.4	BN038060.D	08 Oct 2025 00:17	CCAL failed for Fluoranthene-d10	RC/JU	Ok,M
22	PB169992BL	PB169992BL	BN038061.D	08 Oct 2025 00:54	CCAL failed	RC/JU	Not Ok
23	PB169992BS	PB169992BS	BN038062.D	08 Oct 2025 01:30	CCAL failed	RC/JU	Not Ok
24	PB169992BSD	PB169992BSD	BN038063.D	08 Oct 2025 02:06	CCAL failed	RC/JU	Not Ok
25	SSTDCCC0.4	SSTDCCC0.4EC	BN038064.D	08 Oct 2025 02:43		RC/JU	Ok

M : Manual Integration