

## Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3237

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN038043.D

| Lab Sample ID | Parameter   | Spike | Result | Unit | Rec | RPD | Qual | RPD  | Limits |      | RPD |
|---------------|-------------|-------|--------|------|-----|-----|------|------|--------|------|-----|
|               |             |       |        |      |     |     |      | Qual | Low    | High |     |
| PB169939BS    | 1,4-Dioxane | 0.4   | 0.35   | ug/L | 88  |     |      |      | 70     | 130  |     |