

Initial and Continuing Calibration Verification

Client: Yannuzzi Group, Inc.

SDG No.: Q3238

Project: Edison Yard

RunNo.: LB137364

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.097	0.099	98	90-110	09/30/2025
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90-110	09/30/2025
Sample ID: CCV2 Cyanide	mg/L	0.24	0.25	96	90-110	09/30/2025
Sample ID: CCV3 Cyanide	mg/L	0.25	0.25	100	90-110	09/30/2025