



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Yannuzzi Group, Inc.

SDG No.: Q3238

Contract: YANN01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB51	Mercury	0.076	+/-0.2	U	0.20	CV	09/30/2025	09:49	LB137359

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Yannuzzi Group, Inc.

SDG No.: Q3238

Contract: YANN01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB57	Mercury	0.076	+/-0.2	U	0.20	CV	09/30/2025	09:53	LB137359
CCB58	Mercury	0.076	+/-0.2	U	0.20	CV	09/30/2025	10:21	LB137359
CCB59	Mercury	0.076	+/-0.2	U	0.20	CV	09/30/2025	10:54	LB137359
CCB60	Mercury	0.076	+/-0.2	U	0.20	CV	09/30/2025	11:13	LB137359
CCB61	Mercury	0.076	+/-0.2	U	0.20	CV	09/30/2025	11:40	LB137359

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Yannuzzi Group, Inc.

SDG No.: Q3238

Contract: YANN01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	10/01/2025	12:45	LB137389
	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	12:45	LB137389
	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	12:45	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	12:45	LB137389
	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	12:45	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	12:45	LB137389
	Calcium	234	+/-1000	U	2000	P	10/01/2025	12:45	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	12:45	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	12:45	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	12:45	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	12:45	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	12:45	LB137389
	Magnesium	244	+/-1000	U	2000	P	10/01/2025	12:45	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	12:45	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	12:45	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	12:45	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	12:45	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	12:45	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	12:45	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	12:45	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	12:45	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	12:45	LB137389

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Yannuzzi Group, Inc.

SDG No.: Q3238

Contract: YANN01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	10/01/2025	13:14	LB137389
	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	13:14	LB137389
	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	13:14	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	13:14	LB137389
	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	13:14	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	13:14	LB137389
	Calcium	234	+/-1000	U	2000	P	10/01/2025	13:14	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	13:14	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	13:14	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	13:14	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	13:14	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	13:14	LB137389
	Magnesium	244	+/-1000	U	2000	P	10/01/2025	13:14	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	13:14	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	13:14	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	13:14	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	13:14	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	13:14	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	13:14	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	13:14	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	13:14	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	13:14	LB137389
CCB02	Aluminum	11.9	+/-50	J	100	P	10/01/2025	14:06	LB137389
	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	14:06	LB137389
	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	14:06	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	14:06	LB137389
	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	14:06	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	14:06	LB137389
	Calcium	234	+/-1000	U	2000	P	10/01/2025	14:06	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	14:06	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	14:06	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	14:06	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	14:06	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	14:06	LB137389
	Magnesium	244	+/-1000	U	2000	P	10/01/2025	14:06	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	14:06	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	14:06	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	14:06	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	14:06	LB137389

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Yannuzzi Group, Inc.

SDG No.: Q3238

Contract: YANN01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	10/01/2025	14:06	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	14:06	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	14:06	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	14:06	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	14:06	LB137389
CCB03	Aluminum	26.4	+/-50	J	100	P	10/01/2025	15:04	LB137389
	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	15:04	LB137389
	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	15:04	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	15:04	LB137389
	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	15:04	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	15:04	LB137389
	Calcium	234	+/-1000	U	2000	P	10/01/2025	15:04	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	15:04	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	15:04	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	15:04	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	15:04	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	15:04	LB137389
	Magnesium	244	+/-1000	U	2000	P	10/01/2025	15:04	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	15:04	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	15:04	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	15:04	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	15:04	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	15:04	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	15:04	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	15:04	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	15:04	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	15:04	LB137389
CCB04	Aluminum	11.3	+/-50	U	100	P	10/01/2025	15:55	LB137389
	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	15:55	LB137389
	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	15:55	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	15:55	LB137389
	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	15:55	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	15:55	LB137389
	Calcium	234	+/-1000	U	2000	P	10/01/2025	15:55	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	15:55	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	15:55	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	15:55	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	15:55	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	15:55	LB137389

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Yannuzzi Group, Inc.

SDG No.: Q3238

Contract: YANN01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	10/01/2025	15:55	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	15:55	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	15:55	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	15:55	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	15:55	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	15:55	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	15:55	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	15:55	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	15:55	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	15:55	LB137389
CCB05	Aluminum	11.3	+/-50	U	100	P	10/01/2025	16:46	LB137389
	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	16:46	LB137389
	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	16:46	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	16:46	LB137389
	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	16:46	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	16:46	LB137389
	Calcium	234	+/-1000	U	2000	P	10/01/2025	16:46	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	16:46	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	16:46	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	16:46	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	16:46	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	16:46	LB137389
	Magnesium	244	+/-1000	U	2000	P	10/01/2025	16:46	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	16:46	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	16:46	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	16:46	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	16:46	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	16:46	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	16:46	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	16:46	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	16:46	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	16:46	LB137389
CCB06	Aluminum	11.3	+/-50	U	100	P	10/01/2025	17:41	LB137389
	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	17:41	LB137389
	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	17:41	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	17:41	LB137389
	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	17:41	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	17:41	LB137389
	Calcium	234	+/-1000	U	2000	P	10/01/2025	17:41	LB137389

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Yannuzzi Group, Inc.

SDG No.: Q3238

Contract: YANN01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	17:41	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	17:41	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	17:41	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	17:41	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	17:41	LB137389
	Magnesium	244	+/-1000	U	2000	P	10/01/2025	17:41	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	17:41	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	17:41	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	17:41	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	17:41	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	17:41	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	17:41	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	17:41	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	17:41	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	17:41	LB137389
CCB07	Aluminum	11.4	+/-50	J	100	P	10/01/2025	18:32	LB137389
	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	18:32	LB137389
	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	18:32	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	18:32	LB137389
	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	18:32	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	18:32	LB137389
	Calcium	234	+/-1000	U	2000	P	10/01/2025	18:32	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	18:32	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	18:32	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	18:32	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	18:32	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	18:32	LB137389
	Magnesium	244	+/-1000	U	2000	P	10/01/2025	18:32	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	18:32	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	18:32	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	18:32	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	18:32	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	18:32	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	18:32	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	18:32	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	18:32	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	18:32	LB137389
CCB08	Aluminum	11.3	+/-50	U	100	P	10/01/2025	19:23	LB137389
	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	19:23	LB137389

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Yannuzzi Group, Inc.

SDG No.: Q3238

Contract: YANN01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	19:23	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	19:23	LB137389
	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	19:23	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	19:23	LB137389
	Calcium	234	+/-1000	U	2000	P	10/01/2025	19:23	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	19:23	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	19:23	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	19:23	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	19:23	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	19:23	LB137389
	Magnesium	244	+/-1000	U	2000	P	10/01/2025	19:23	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	19:23	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	19:23	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	19:23	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	19:23	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	19:23	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	19:23	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	19:23	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	19:23	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	19:23	LB137389