

Continuing Calibration Report for SequenceID : FD100225AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049811.D

Aromatic C10-C12	4706995.000	40.000	4.389	6.131	117674.875	122789.439	4.165
Aromatic C12-C16	8303318.000	60.000	6.132	8.759	138388.633	143358.701	3.467
Aromatic C16-C21	10980590.000	80.000	8.760	13.049	137257.375	142171.561	3.457
Aromatic C21-C36	20756349.000	180.000	13.050	18.468	115313.050	123821.788	6.872
Aromatic EPH	44747252.000	360.000	4.389	18.468	124297.922	131040.962	5.146