

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** YANN01
Lab Code: ACE **SDG NO.:** Q3238
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075434.D **Time Analyzed:** 08:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.807	5.708	5.908	530.620	500.000	6.1
Aroclor-1016-2	5.828	5.729	5.929	525.210	500.000	5.0
Aroclor-1016-3	5.890	5.792	5.992	537.720	500.000	7.5
Aroclor-1016-4	5.988	5.889	6.089	583.920	500.000	16.8
Aroclor-1016-5	6.280	6.181	6.381	571.420	500.000	14.3
Aroclor-1260-1	7.397	7.299	7.499	522.290	500.000	4.5
Aroclor-1260-2	7.650	7.551	7.751	514.200	500.000	2.8
Aroclor-1260-3	8.008	7.909	8.109	563.460	500.000	12.7
Aroclor-1260-4	8.237	8.137	8.337	588.400	500.000	17.7
Aroclor-1260-5	8.562	8.464	8.664	573.040	500.000	14.6
Decachlorobiphenyl	10.432	10.334	10.534	54.160	50.000	8.3
Tetrachloro-m-xylene	4.655	4.555	4.755	52.100	50.000	4.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** YANN01
Lab Code: ACE **SDG NO.:** Q3238
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075434.D **Time Analyzed:** 08:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.890	4.793	4.993	546.990	500.000	9.4
Aroclor-1016-2	4.949	4.851	5.051	555.350	500.000	11.1
Aroclor-1016-3	5.068	4.971	5.171	493.970	500.000	-1.2
Aroclor-1016-4	5.108	5.011	5.211	555.270	500.000	11.1
Aroclor-1016-5	5.323	5.225	5.425	535.230	500.000	7.0
Aroclor-1260-1	6.351	6.255	6.455	522.520	500.000	4.5
Aroclor-1260-2	6.537	6.441	6.641	513.690	500.000	2.7
Aroclor-1260-3	6.691	6.595	6.795	510.050	500.000	2.0
Aroclor-1260-4	7.161	7.065	7.265	482.390	500.000	-3.5
Aroclor-1260-5	7.400	7.304	7.504	501.480	500.000	0.3
Decachlorobiphenyl	8.798	8.703	8.903	62.930	50.000	25.9
Tetrachloro-m-xylene	3.794	3.696	3.896	56.350	50.000	12.7