

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: YANN01

Lab Code: ACE

SDG NO.: Q3238

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097468.D Date Analyzed: 10/01/2025

Lab Sample No.(PEM): PEM Time Analyzed: 09:45

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.997	8.900	9.100	20.340	20.000	1.7
Tetrachloro-m-xylene	3.525	3.470	3.580	18.080	20.000	-9.6
alpha-BHC	3.972	3.920	4.020	8.300	10.000	-17.0
beta-BHC	4.487	4.440	4.540	9.070	10.000	-9.3
gamma-BHC (Lindane)	4.300	4.250	4.350	8.550	10.000	-14.5
Endrin	6.529	6.460	6.600	45.140	50.000	-9.7
4,4'-DDT	6.976	6.910	7.050	77.440	100.000	-22.6
Methoxychlor	7.449	7.380	7.520	191.110	250.000	-23.6

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/18/2025 09/18/2025

Client Sample No. (PEM): PEM - PL097468.D Date Analyzed: 10/01/2025

Lab Sample No.(PEM): PEM Time Analyzed: 09:45

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.981	7.880	8.080	17.000	20.000	-15.0
Tetrachloro-m-xylene	2.822	2.770	2.870	16.820	20.000	-15.9
alpha-BHC	3.327	3.280	3.380	7.770	10.000	-22.3
beta-BHC	3.955	3.900	4.010	8.430	10.000	-15.7
gamma-BHC (Lindane)	3.659	3.610	3.710	7.840	10.000	-21.6
Endrin	5.700	5.630	5.770	42.000	50.000	-16.0
4,4'-DDT	6.099	6.030	6.170	70.940	100.000	-29.1
Methoxychlor	6.671	6.600	6.740	156.050	250.000	-37.6