

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097271.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 17:06
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 04:44:06 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 04:43:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.527	2.823	455.4E6	654.7E6	101.994	101.503
28)	SA Decachlor...	9.001	7.984	315.2E6	536.1E6	98.872	98.183
Target Compounds							
2)	A alpha-BHC	3.974	3.329	743.3E6	1044.5E6	105.888	104.303
3)	MA gamma-BHC...	4.302	3.660	694.0E6	963.3E6	105.254	103.575
4)	MA Heptachlor	4.893	4.008	680.9E6	929.9E6	103.043	100.901
5)	MB Aldrin	5.232	4.290	655.2E6	886.1E6	104.089	102.301
6)	B beta-BHC	4.489	3.956	269.2E6	397.0E6	100.455	100.557
7)	B delta-BHC	4.735	4.189	640.9E6	958.6E6	104.552	103.832
8)	B Heptachlo...	5.652	4.793	576.4E6	782.6E6	100.128	99.922
9)	A Endosulfan I	6.033	5.163	524.9E6	738.7E6	102.490	99.419
10)	B gamma-Chl...	5.905	5.044	579.3E6	855.4E6	103.252	99.860
11)	B alpha-Chl...	5.986	5.108	572.8E6	827.5E6	102.408	100.361
12)	B 4,4'-DDE	6.156	5.297	496.4E6	791.1E6	104.790	102.753
13)	MA Dieldrin	6.306	5.428	566.8E6	825.1E6	103.988	101.465
14)	MA Endrin	6.532	5.702	477.8E6	745.0E6	103.448	101.517
15)	B Endosulfa...	6.745	5.994	469.5E6	697.8E6	98.323	100.299
16)	A 4,4'-DDD	6.665	5.850	389.1E6	658.9E6	103.408	103.640
17)	MA 4,4'-DDT	6.979	6.102	440.3E6	719.0E6	103.710	103.115
18)	B Endrin al...	6.874	6.172	303.6E6	476.2E6	99.226	96.896
19)	B Endosulfa...	7.108	6.396	408.4E6	666.5E6	101.600	99.231
20)	A Methoxychlor	7.452	6.674	205.3E6	358.1E6	99.532	96.808
21)	B Endrin ke...	7.587	6.900	429.6E6	722.3E6	101.695	97.725
22)	Mirex	8.065	7.089	330.8E6	552.6E6	97.189	96.302

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097271.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 17:06
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 04:44:06 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 04:43:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

