

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
 Data File : PL097272.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2025 17:19
 Operator : AR\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 04:44:36 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 04:43:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.528	2.824	334.5E6	484.3E6	74.906	75.085
2)	SA Decachlor...	9.002	7.985	234.1E6	398.1E6	73.453	72.915
Target Compounds							
2)	A alpha-BHC	3.975	3.329	536.3E6	757.5E6	76.392	75.640
3)	MA gamma-BHC...	4.302	3.661	502.5E6	703.0E6	76.204	75.591
4)	MA Heptachlor	4.893	4.009	496.6E6	687.2E6	75.148	74.565
5)	MB Aldrin	5.232	4.291	477.7E6	650.5E6	75.880	75.100
6)	B beta-BHC	4.490	3.957	198.1E6	292.5E6	73.934	74.080
7)	B delta-BHC	4.735	4.190	463.0E6	697.0E6	75.527	75.496
8)	B Heptachlo...	5.652	4.793	424.1E6	581.2E6	73.678	74.211
9)	A Endosulfan I	6.034	5.163	383.9E6	550.9E6	74.971	74.138
10)	B gamma-Chl...	5.906	5.045	423.9E6	636.8E6	75.554	74.339
11)	B alpha-Chl...	5.986	5.109	419.8E6	614.2E6	75.053	74.493
12)	B 4,4'-DDE	6.156	5.297	361.5E6	581.0E6	76.313	75.459
13)	MA Dieldrin	6.306	5.428	412.9E6	609.0E6	75.745	74.885
14)	MA Endrin	6.533	5.703	348.1E6	550.2E6	75.368	74.968
15)	B Endosulfa...	6.745	5.995	347.8E6	518.4E6	72.847	74.509
16)	A 4,4'-DDD	6.665	5.850	284.1E6	483.2E6	75.512	76.000
17)	MA 4,4'-DDT	6.980	6.102	320.6E6	526.7E6	75.521	75.530
18)	B Endrin al...	6.875	6.173	224.8E6	356.8E6	73.496	72.597
19)	B Endosulfa...	7.108	6.396	299.3E6	496.8E6	74.448	73.965
20)	A Methoxychlor	7.453	6.674	152.1E6	268.7E6	73.718	72.634
21)	B Endrin ke...	7.588	6.900	314.3E6	536.5E6	74.403	72.588
22)	Mirex	8.066	7.090	246.6E6	415.4E6	72.461	72.394

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092025\
Data File : PL097272.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2025 17:19
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 04:44:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 04:43:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

