

Surrogate Summary

SDG No.: **Q3238**

Client: **Yannuzzi Group, Inc.**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PL097268.D	PIBLK-PL097268.D	Decachlorobiphen	1	20	17.3	87		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	15.6	78		30 (61)	150 (148)
		Decachlorobiphen	2	20	17.1	86		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	16.0	80		30 (61)	150 (148)
I.BLK-PL097467.D	PIBLK-PL097467.D	Decachlorobiphen	1	20	24.8	124		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	21.9	109		30 (61)	150 (148)
		Decachlorobiphen	2	20	20.3	102		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	20.2	101		30 (61)	150 (148)
PB169908BL	PB169908BL	Decachlorobiphen	1	20	23.7	118		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	20.7	103		30 (19)	150 (148)
		Decachlorobiphen	2	20	20.2	101		30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	19.6	98		30 (19)	150 (148)
PB169908BS	PB169908BS	Decachlorobiphen	1	20	20.6	103		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	17.9	90		30 (19)	150 (148)
		Decachlorobiphen	2	20	17.7	88		30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	16.7	83		30 (19)	150 (148)
Q3238-01	RCA	Decachlorobiphen	1	20	14.6	73		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	13.2	66		30 (19)	150 (148)
		Decachlorobiphen	2	20	12.9	64		30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	12.5	63		30 (19)	150 (148)
Q3245-01MS	13-1MS	Decachlorobiphen	1	20	13.8	69		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	13.8	69		30 (19)	150 (148)
		Decachlorobiphen	2	20	11.9	59		30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	14.0	70		30 (19)	150 (148)
Q3245-01MSD	13-1MSD	Decachlorobiphen	1	20	14.3	71		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	14.1	71		30 (19)	150 (148)
		Decachlorobiphen	2	20	12.4	62		30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	14.2	71		30 (19)	150 (148)
I.BLK-PL097482.D	PIBLK-PL097482.D	Decachlorobiphen	1	20	24.7	124		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	21.1	106		30 (61)	150 (148)
		Decachlorobiphen	2	20	21.5	108		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	20.6	103		30 (61)	150 (148)