

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Alliane Contract: YANN01
ProjectID: Edison Yard
Lab Code: ACE SDG No.: Q3238
DataFile: FF016493.D Analyst Name: YP\AJ Analyst Date: 10-01-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
850	100251627	117943	126783	6.973