

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Alliane Contract: YANN01
ProjectID: Edison Yard
Lab Code: ACE SDG No.: Q3238
DataFile: FF016506.D Analyst Name: YP\AJ Analyst Date: 10-01-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
850	94216394	110843	126783	12.573