

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3238

Client ID : SSTDCCC040

Date Analyzed: 09/30/2025

Lab File ID: BF143830.D

Time Analyzed: 13:36

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	130920	6.886	478225	8.17	247405	9.93
UPPER LIMIT	261840	7.386	956450	8.669	494810	10.427
LOWER LIMIT	65460	6.386	239113	7.669	123703	9.427
EPA SAMPLE NO.						
01 12-8MS	127168	6.89	463068	8.17	231503	9.92
02 12-8MSD	124507	6.89	452928	8.17	222658	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.