

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3238

Client ID : SSTDCCC040

Date Analyzed: 09/30/2025

Lab File ID: BP025892.D

Time Analyzed: 15:26

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	161338	7.737	638191	10.50	423574	14.34
UPPER LIMIT	322676	8.237	1276380	11.002	847148	14.843
LOWER LIMIT	80669	7.237	319096	10.002	211787	13.843
EPA SAMPLE NO.						
01 PB169909BL	118992	7.74	457573	10.50	308102	14.35
02 PB169909BS	125775	7.74	504241	10.50	321119	14.35
03 RCA	115585	7.74	449301	10.51	307634	14.35

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.