

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3238

Client ID: SSTDCCC040

Date Analyzed: 09/30/2025

Lab File ID: BP025892.D

Time Analyzed: 15:26

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	884109	17.148	926260	21.595	1078930	24.924
UPPER LIMIT	1768220	17.648	1852520	22.095	2157860	25.424
LOWER LIMIT	442055	16.648	463130	21.095	539465	24.424
EPA SAMPLE NO.						
01 PB169909BL	685664	17.15	824101	21.60	924310	24.94
02 PB169909BS	671502	17.15	743351	21.59	855731	24.94
03 RCA	669139	17.15	816530	21.60	942797	24.96

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.