

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: YANN01
 Lab Code: ACE SDG No.: Q3238
 Instrument ID: MSVOA_W Calibration Date/Time: 09/30/2025 12:54
 Lab File ID: VW032298.D Init. Calib. Date(s): 09/25/2025 09/25/2025
 Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:45 13:08
 GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	49.82	0.3639984	20