

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW092525

Review By	Maresh Dadoda	Review On	9/29/2025 1:07:01 AM
Supervise By	Semsettin Yesilyurt	Supervise On	9/29/2025 1:07:30 AM
SubDirectory	VW092525	HP Acquire Method	MSVOA_W HP Processing Method 82w092525s.m

STD. NAME	STD REF.#
Tune/Reschk	VP135637
Initial Calibration Stds	VP135638,VP135639,VP135641,VP135642,VP135643,VP135644
CCC	
Internal Standard/PEM	VP133934
ICV/I.BLK	VP135645
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VW032257.D	25 Sep 2025 10:01		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VW032258.D	25 Sep 2025 10:45	LR-20	SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VW032259.D	25 Sep 2025 11:30		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VW032260.D	25 Sep 2025 11:52		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VW032261.D	25 Sep 2025 12:24		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VW032262.D	25 Sep 2025 12:46		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VW032263.D	25 Sep 2025 13:08		SY/MD	Ok,M
8	VIBLK	VIBLK	VW032264.D	25 Sep 2025 13:46		SY/MD	Ok
9	VSTDICV050	ICVVW092525	VW032265.D	25 Sep 2025 14:41		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW092925

Review By	Semsettin Yesilyurt	Review On	10/3/2025 2:29:37 AM
Supervise By	Maresh Dadoda	Supervise On	10/3/2025 2:32:52 AM
SubDirectory	VW092925	HP Acquire Method	MSVOA_W HP Processing Method 82w092525s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135691
CCC	VP135692,VP135693
Internal Standard/PEM	VP133934
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VW032287.D	29 Sep 2025 08:46		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VW032288.D	29 Sep 2025 09:18		SY/MD	Ok,M
3	VW0929SBL01	VW0929SBL01	VW032289.D	29 Sep 2025 10:18		SY/MD	Ok
4	VW0929SBS01	VW0929SBS01	VW032290.D	29 Sep 2025 10:46		SY/MD	Ok,M
5	VW0929SBSD01	VW0929SBSD01	VW032291.D	29 Sep 2025 11:09		SY/MD	Ok,M
6	Q3221-01RE	HD-01-092625RE	VW032292.D	29 Sep 2025 11:52	vial-B Internal Standard Fail	SY/MD	Confirms
7	Q3220-01RE	OR-03-092625RE	VW032293.D	29 Sep 2025 12:15	vial-B Internal Standard Fail	SY/MD	Confirms
8	Q3238-01	RCA	VW032294.D	29 Sep 2025 14:16	vial-A Surrogate Fail	SY/MD	ReRun
9	VIBLK	VIBLK	VW032295.D	29 Sep 2025 14:38		SY/MD	Ok
10	VSTDCCC050	VSTDCCC050EC	VW032296.D	29 Sep 2025 15:00		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW093025

Review By	Mahesh Dadoda	Review On	10/3/2025 2:25:38 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	10/3/2025 2:30:21 AM		
SubDirectory	VW093025	HP Acquire Method	MSVOA_W	HP Processing Method	82w092525s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP135700			
CCC		VP135701,VP135702			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VW032297.D	30 Sep 2025 10:43		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VW032298.D	30 Sep 2025 12:54		SY/MD	Ok,M
3	VW0930SBL01	VW0930SBL01	VW032299.D	30 Sep 2025 13:25		SY/MD	Ok
4	VW0930SBS01	VW0930SBS01	VW032300.D	30 Sep 2025 13:51		SY/MD	Ok,M
5	VW0930SBSD01	VW0930SBSD01	VW032301.D	30 Sep 2025 14:14		SY/MD	Ok,M
6	Q3238-01RE	RCARE	VW032302.D	30 Sep 2025 14:52	vial-B Surrogate fail	SY/MD	Confirms
7	Q3150-04	Mid Site Grid 5-7	VW032303.D	30 Sep 2025 15:14	vial-A Internal Standard Fail	SY/MD	ReRun
8	Q3253-01	ARS20-0009	VW032304.D	30 Sep 2025 15:36	vial-A	SY/MD	Ok
9	Q3245-02	13-1	VW032305.D	30 Sep 2025 15:57	vial-A	SY/MD	Ok
10	Q3245-10	12-8	VW032306.D	30 Sep 2025 16:19	vial-A	SY/MD	Ok
11	Q3245-18	12-1	VW032307.D	30 Sep 2025 16:41	vial-A	SY/MD	Ok
12	Q3240-01	OK-02-092925	VW032308.D	30 Sep 2025 17:04	vial-A	SY/MD	Ok
13	Q3256-03	SOIL-PILE-1-WC-VOC	VW032309.D	30 Sep 2025 17:26	vial-A	SY/MD	Ok
14	Q3256-07	SOIL-PILE-2-WC-VOC	VW032310.D	30 Sep 2025 17:48	vial-A	SY/MD	Ok
15	VIBLK	VIBLK	VW032311.D	30 Sep 2025 18:10		SY/MD	Ok

M : Manual Integration