

Surrogate Summary

SDG No.: **Q3238**

Client: **Yannuzzi Group, Inc.**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3238-01	RCA	1,2-Dichloroethane-d4	50	60.0	120		70 (63)	130 (155)
		Dibromofluoromethane	50	26.8	54	*	70 (70)	130 (134)
		Toluene-d8	50	51.0	102		70 (74)	130 (123)
		4-Bromofluorobenzene	50	48.2	96		70 (17)	130 (146)
Q3238-01RE	RCARE	1,2-Dichloroethane-d4	50	60.8	122		70 (63)	130 (155)
		Dibromofluoromethane	50	23.5	47	*	70 (70)	130 (134)
		Toluene-d8	50	48.9	98		70 (74)	130 (123)
		4-Bromofluorobenzene	50	45.7	91		70 (17)	130 (146)
VW0929SBL01	VW0929SBL01	1,2-Dichloroethane-d4	50	62.9	126		70 (63)	130 (155)
		Dibromofluoromethane	50	50.9	102		70 (70)	130 (134)
		Toluene-d8	50	51.9	104		70 (74)	130 (123)
		4-Bromofluorobenzene	50	49.6	99		70 (17)	130 (146)
VW0929SBS01	VW0929SBS01	1,2-Dichloroethane-d4	50	49.1	98		70 (63)	130 (155)
		Dibromofluoromethane	50	48.6	97		70 (70)	130 (134)
		Toluene-d8	50	50.3	101		70 (74)	130 (123)
		4-Bromofluorobenzene	50	51.6	103		70 (17)	130 (146)
VW0929SBSD0	VW0929SBSD01	1,2-Dichloroethane-d4	50	48.9	98		70 (63)	130 (155)
		Dibromofluoromethane	50	47.0	94		70 (70)	130 (134)
		Toluene-d8	50	48.8	97		70 (74)	130 (123)
		4-Bromofluorobenzene	50	48.8	98		70 (17)	130 (146)
VW0930SBL01	VW0930SBL01	1,2-Dichloroethane-d4	50	57.0	114		70 (63)	130 (155)
		Dibromofluoromethane	50	50.2	100		70 (70)	130 (134)
		Toluene-d8	50	51.1	102		70 (74)	130 (123)
		4-Bromofluorobenzene	50	46.1	92		70 (17)	130 (146)
VW0930SBS01	VW0930SBS01	1,2-Dichloroethane-d4	50	49.9	100		70 (63)	130 (155)
		Dibromofluoromethane	50	46.7	93		70 (70)	130 (134)
		Toluene-d8	50	46.6	93		70 (74)	130 (123)
		4-Bromofluorobenzene	50	46.6	93		70 (17)	130 (146)

